

305.1 - Surface Flammability (sheet form)

This SRM is for checking the operation of radiant panel test equipment in accordance with the procedures outlined in [ASTM E 162](#)–78.

Technical Contact: william.grosshandler@nist.gov

PLEASE NOTE: The tables are presented to facilitate comparisons among a family of materials to help customers select the best SRM for their needs. For specific values and uncertainties, the certificate is the only official source.

SRM	Description	Unit Size	Certification
1002d	Hard Board (Surface Flammability)	set (4)	Flame Spread Index, I = 203 Heat Evolution Factor, Q = 42.0